

Document Id:	PD-702046 Rev:1
State:	Released
Created by:	Zahrah Chaudhary
Title:	Protective Aprons
Page 1(2)	

We, Mölnlycke Health Care AB, Gamlestadsvägen 3C, Box 13080, SE-402 52 Göteborg, Sweden being exclusively responsible manufacturer for conformity of the device/s; declare that the devices listed in the attached schedule are in conformity with the provisions of the Medical Device Regulation 2017/745, concerning medical devices.

Other Union Legislation applicable: Not applicable

Trade name / Product name:	Protective Aprons
Product Classification:	MDR Class I ns
MDR Classification Rule:	1
Sterility:	Non-sterile
Measuring Function:	No

This declaration is supported by a conformity assessment procedure in accordance with Annex/es: IV

Common Specification: No CS is applicable

Certificate number:	Not Applicable	
Issued by:	Not Applicable	ld No:
	(Notified Body Name)	
Note. For non-sterile, non-measuring class I products there is no Notified Body certificate		

Mölnlycke Health Care issues this declaration in recognition of all applicable harmonised standards.

Signed for and on behalf of Mölnlycke Health Care

Place of Issue: Göteborg, Sweden

2021-06-23 Date:

Function:

Christina Lewing Name:

Signature:

Clumber of the Clumbe



Generic product information		
Intended Purpose:	The product should be worn in situations where there is a risk that infectious agents can be transferred throughout the hospital. Secondly the product also protects the clothes of the user from different kind of fluids.	
Basic UDI-DI:	73324300000000097KH	
GMDN Code:	40511 Hygiene apron, single-use	

Product References Covered by this Declaration:	Product Descriptor:
613000	APRON
613100	Protective Apron