

Title: Filtering Half Mask

Page 1(2)

We, Mölnlycke Health Care AB, Gamlestadvägen 3C, Box 13080, SE-402 52 Göteborg, Sweden being exclusively responsible manufacturer for conformity of the device/s; declare that the devices listed in the attached schedule are in conformity with the provisions of the Council Directive 93/42/EEC of 14 June 1993, as amended by 2007/47/EEC, concerning medical devices, the Medical Devices Act SFS 1993:584 and the Swedish Medical Product Agency regulations and guidelines: Medical Devices, LVFS 2003:11, as amended by LVFS 2009:18.

Trade name/ Product name:	Filtering Half Mask
------------------------------	----------------------------

Product classification: **I**
Sterility Status: **Non-sterile**
Measuring function: **No**

This declaration is supported by a conformity assessment procedure in accordance with	
Annex/es:	VII

Certificate number:	Not Applicable
Issued by:	Not Applicable
For non sterile, non-measuring Class I products, no certificate is issued by a Notified Body.	

Mölnlycke Health Care issues this declaration in recognition of all applicable harmonised standards.

Signed for and on behalf of Mölnlycke Health Care

Date: **2020-07-10** Function: **Regulatory Affairs Director**

Name: **Christina Lewing**

Signature:



Product(s) covered by this declaration:

Product Reference:	Product Descriptor:	GMDN Code:
42902	FILTERING HALF MASK	57794 Surgical/ medical respirator, non-antimicrobia I, single-use
42904	FILTERING HALF MASK	57794 Surgical/ medical respirator, non-antimicrobia I, single-use